



## NUCLEAR SUBSTANCE ROOM DECOMMISSIONING FORM

The permit holder shall ensure that prior to decommissioning any area, room or enclosure where the permitted activity has been conducted: non-fixed contamination does not exceed  $0.3 \text{ Bq/cm}^2$

for all class A radionuclides and  $0.6 \text{ Bq/cm}^2$  for class B radionuclides. For information on the limits for class A and B radionuclides, see 10 CFR 20.171 and 10 CFR 20.172.

---

---

